

Phacic Eyes - Estimated mean IOP					
Time	Catheter	n	Mean	Std. Error	95% Confidence interval
A) Baseline	iTRACK™ 250A	43	22,30	0,99	20,3501 to 24,25
	Glaucolight	37	23,08	1,07	20,9765 to 25,18
B) 1st Day	iTRACK™ 250A	43	11,72	0,99	9,7687 to 13,67
	Glaucolight	37	13,64	1,07	11,5441 to 15,75
C) 1-3 Months	iTRACK™ 250A	41	14,70	1,01	12,7081 to 16,70
	Glaucolight	34	17,23	1,11	15,0399 to 19,43
D) 3-6 Months	iTRACK™ 250A	38	15,55	1,05	13,4759 to 17,62
	Glaucolight	24	17,37	1,32	14,7619 to 19,98
E) 9-12 Months	iTRACK™ 250A	31	15,80	1,16	13,5072 to 18,10
	Glaucolight	23	16,26	1,35	13,5916 to 18,93
F) 12-18 Months	iTRACK™ 250A	14	16,64	1,73	13,2215 to 20,06
	Glaucolight	7	16,57	2,46	11,7329 to 21,41
Pseudophacic eyes – Estimated mean IOP					
Time	Catheter	n	Mean	Std. Error	95% Confidence interval
A) Baseline	iTRACK™ 250A	13	20,92	1,59	17,73 to 24,10
	Glaucolight	9	16,55	1,92	12,72 to 20,38
B) 1st Day	iTRACK™ 250A	13	12,76	1,59	9,58 to 15,95
	Glaucolight	9	11,00	1,92	7,17 to 14,82
C) 1-3 Months	iTRACK™ 250A	10	16,80	1,82	13,17 to 20,42
	Glaucolight	7	16,00	2,18	11,66 to 20,33
D) 3-6 Months	iTRACK™ 250A	9	14,11	1,92	10,28 to 17,93
	Glaucolight	7	14,00	2,18	9,66 to 18,33
E) 9-12 Months	iTRACK™ 250A	7	14,71	2,18	10,37 to 19,05
	Glaucolight	4	15,25	2,88	9,51 to 20,98
F) 12-18 Months	iTRACK™ 250A	4	13,50	2,88	7,76 to 19,23
	Glaucolight	1	11,00	5,76	-0,47 to 22,47

Table 7. Estimated marginal means of IOP following Canaloplasty ab externo between iTRACK™ 250A and Glaucolight, in phacic and pseudophacic eyes in 2 Ways ANOVA.